

O2k-FluoRespirometry: Amplex UltraRed using freeze-dried baker's yeast

Fontana-Ayoub M¹, Fasching M¹, Eigentler A², Laner V¹, Gnaiger E^{1,2}

¹Oroboros Instruments

High-Resolution FluoRespirometry
Schöpfstr 18, A-6020 Innsbruck, Austria
Email: instruments@orooboros.at
www.orooboros.at

²D. Swarovski Research Lab
Dept of Visceral, Transplant and Thoracic Surgery
Medical Univ Innsbruck, 6020 Innsbruck, Austria
www.mitofit.org

Contents

1. Introduction	1
2. Chemicals	2
3. Experimental setup with the Oxygraph-2k	2
4. Experimental protocol	3
5. Results and discussion	3
6. References	4
Acknowledgements	4
Author contributions and publication versions	4

1. Introduction

Baker's yeast is an anhydrobiotic organism which is well known for its persistence without water for decades. When rehydrated, it can rapidly restore active metabolism within minutes [1]. In addition, dried baker's yeast has a high level of viability when rehydrated at 30 to 40 °C [1,2]. Commercially available freeze-dried baker's yeast is rehydrated in Na-Phosphate buffer (1 mL). Pipetting up and down 20 times and 10 min on a shaker are necessary to obtain a homogenous cell suspension. Subsequently, another 10 times carefully pipetting up and down are required, immediately prior to adding a subsample of 20 µL into the 2 mL O2k-chamber (37 °C). Since the cells sediment and clump rapidly, it is necessary to pipette up and down the stock cell suspension again 10 times prior to addition into the next chamber.

2. Chemicals

Respiration medium: MiR05-Kit (Na-PB 50 mM, pH 7.1)

Yeast stock suspension (20 mg/mL):

- Dissolve 20 mg baker's yeast (active dry yeast granula) in 1 mL warm respiration medium [3].
- Pipette up and down slowly 20 times to obtain a homogenous cell suspension.
- Pipette up and down more quickly another 10 times and fill the syringe. Titrate 20 (100) μ L yeast stock suspension into each 2 mL-chamber (the final concentration in the chamber = 0.2 (1.0) mg/mL; *in the present example 1.5 mg/mL was used*).
- Repeat pipetting up and down 10 times before each addition to the O2k-chamber.

H₂O₂ calibration solution (40 μ M stock):

Commercial solution (Sigma-Aldrich 323381 - 25 mL hydrogen peroxide solution 3 wt.%) = 880 mM; prepare fresh:

- H₂O₂ dilution 1 (1:88): add 284 μ L of a commercial solution to 20 mL of 10 μ M HCl; fill it up to a final volume of 25 mL with 10 μ M HCl, to obtain a 10 mM H₂O₂ solution.
- H₂O₂ dilution 2 (1:125): dilute 200 μ L of dilution 1 with 10 μ M HCl to a volume of 50 mL to obtain the H₂O₂ stock solution of **40 μ M**.

Amplex[®] UltraRed (Invitrogen A 36006): Storage solution=stock solution 10 mM AmR in DMSO. For details see:

» www.bioblast.at/index.php/Amplex_red

Horse radish peroxidase (Sigma-Aldrich P 8250 - 5 kU): Stock solution 500 U HRP/mL MiR05 or MiR05Cr; can be used as storage solution at -20 °C.

CCCP (Sigma-Aldrich C2759): 1 mM stock solution (1.02 mg in 5 mL DMSO). Intact yeast requires much higher concentrations of uncoupler compared to mammalian cells or mitochondrial preparations.

FCCP (Sigma-Aldrich C2920): 1 mM stock solution; now replaced by CCCP www.bioblast.at/index.php/Bioblast_alert#Bioblast_alert_2013.2802.29:2013-08-08

The solvent ethanol for FCCP has the disadvantage that it stimulates respiration as a substrate oxidized by yeast. This is avoided by DMSO as a solvent for CCCP.

3. Experimental setup with the Oxygraph-2k

Temperature: 37 °C

Gain for OroboPOS: 1

Polarisation voltage for OroboPOS: 800 mV

Gain for amperometric (Amp) sensor: 1000

Polarisation voltage for Amp sensor: 100 mV

Data recording interval: 2 s

Oxygen solubility factor of 50 mM Na-PB: $F_M = 0.95$ (for MiR05: 0.92)

4. Experimental protocol

Titrations (final conc.)	Event name	Additions [μ l]
Yeast (0.2 to 1.0 mg/mL)*	Yeast	20 to 100
Amplex [®] UltraRed (10 μ M)	AmR	2
Horse radish peroxidase (1 U/mL)	HRP	4
SOD (5 U/mL)	SOD	Variable
H ₂ O ₂ (0.1 μ M steps, total 0.2 μ M)	H ₂ O ₂	5 + 5 (two steps)
Glucose (20 mM)	Glc (2 M stock)	20
EtOH	EtOH (pure)	40
CCCP (5 μ M steps)**	U	10+10+10+10+..
H ₂ O ₂ (0.1 μ M steps, total 0.2 μ M)	H ₂ O ₂	5 + 5 (two steps)

Reoxygenations were performed with gas injections of pure oxygen.

*Original experiments: 1.5 mg/mL; **FCCP was used, now replaced by CCCP.

5. Results and discussion

In the study presented in Figure 1 (conditions differed from the above recommendations [4]), respiration and H₂O₂ production of yeast (1.5 mg/mL) was studied under hypoxic and hyperoxic conditions, and after stimulation of respiration by glucose. Extracellular H₂O₂-production by intact yeast cells was a pronounced function of environmental oxygen concentration. Oxygen kinetics of respiration (Fig. 1B) resembled closely but not fully that of isolated mitochondria [5], indicating that extracellular oxygen levels provide a close approximation of intracellular conditions, and intracellular oxygen gradients are small. Stimulation of respiration was pronounced with the extracellular substrates glucose and ethanol, and with uncoupler titrations (Fig. 2). In contrast, these additions exerted a minor effect on extracellular H₂O₂ production (Fig. 2).

H₂O₂-Production as a Function of [O₂] in Intact Yeast Cells




Figure 1. O₂k-traces of oxygen concentration (blue plot), oxygen flux (red) and hydrogen peroxide production (green) as a function of time (top) and oxygen concentration (bottom).

Extracellular H₂O₂ production was a linear function of oxygen concentration in the hyperbaric and normoxic range [6]. H₂O₂

production, however, declined steeply in the microoxic range, when the critical oxygen level for respiration was reached (Fig. 1).

