

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

**CB TEST CERTIFICATE**

Product

Laboratory equipment

Name and address of the applicant

Oroboros Instruments GmbH  
Schoepfstrasse 18, 6020 Innsbruck  
AUSTRIA

Name and address of the manufacturer

Oroboros Instruments GmbH  
Schoepfstrasse 18, 6020 Innsbruck  
AUSTRIA

Name and address of the factory

WGT Elektronik GmbH & CoKG  
Rettenbergstraße 30a, 6114 Kolsass  
AUSTRIA

*Note: When more than one factory, please report on page 2*

Additional Information on page 2

Ratings and principal characteristics

AC 100 - 240 V; 50-60 Hz; 0.45 A or 0.6 A (for external power supply) DC 5 V; 1.5 A (for equipment)

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Oxia

Additional information (if necessary may also be reported on page 2)

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010  
IEC 61010-1:2010/AMD1:2016

As shown in the Test Report Ref. No. which forms part of this Certificate

295610-TL4-1

This CB Test Certificate is issued by the National Certification Body

**VDE** Prüf- und Zertifizierungsinstitut GmbH  
**VDE** Testing and Certification Institute  
Zertifizierungsstelle und internat. Angelegenheiten  
Certification Body and internat. Affairs

Date: 2022-10-10

Signature:

B. Megerle